

BIOMEDICAL SYSTEMS

2020, January 24

Certificate of Analysis:

Gram Reagent C: Crystal Violet, 500mL

Catalog Number:	SS-041C-EU
Lot:	160552
Expiration Date:	2021, June 17

Characteristics	Specifications	Results
Appearance	Dark purple solution	Conforms
Optical density at 590 nm	0.700 - 0.820	0.800

Authorized by:

01/27/2020 Justin Allen

Quality Control Supervisor

Date