

BIOMEDICAL SYSTEMS 370 West 1700 South Logan, Utah 84321 USA 435 752 6011 800 453 2725 fax 435 752 4127 web www.elitechgroup.com

2020, March 06

Certificate of Analysis:

Gram Reagent C: Crystal Violet, 500mL

Catalog Number:

SS-041C-EU

Lot:

J09209

Expiration Date:

2021, August 20

Characteristics	Specifications	Results
Appearance	Dark purple solution	Conforms
Optical density at 590 nm	0.700 - 0.820	0.820

Authorized by:

Justin Allen

Quality Control Supervisor

Date