

BIOMEDICAL SYSTEMS 370 West 1700 South Logan, Utah 84321 USA 435 752 6011 800 453 2725 fax 435 752 4127 web www.elltechgroup.com

2020, June 09

Certificate of Analysis:

Gram Reagent C: Crystal Violet, 500mL

Catalog Number: SS-041C-EU Lot: J63920

Expiration Date: 2021, October 15

Characteristics	Specifications	Results
Appearance	Dark purple solution	Conforms
Optical density at 590 nm	0.700 - 0.820	0.760

Justin Allen Date

Quality Control Supervisor

File CAJ63920-EU.docx Page 1 of 1 © ELITechGroup Inc.