



BIOMEDICAL SYSTEMS
370 West 1700 South
Logan, Utah 84321 USA
435 752 6011
800 453 2725
fax 435 752 4127
web www.elitechgroup.com

2021, July 07

Certificate of Analysis:

Gram Reagent C: Crystal Violet, 500mL

Catalog Number: SS-041C-EU
Lot: J02744
Expiration Date: 2022, April 13

Characteristics	Specifications	Results
Appearance	Dark purple solution	Conforms
Optical density at 590 nm	0.700 – 0.820	0.810

Authorized by:

A handwritten signature in blue ink, followed by the date "2021-07-08" written in the same ink.

Whitney Lewis, Director of Quality
