



BIOMEDICAL SYSTEMS

370 West 1700 South
Logan, Utah 84321 USA
435 752 6011
800 453 2725
fax 435 752 4127
web www.elitechgroup.com

2021, June 17

Certificate of Analysis:

Gram Reagent C: Crystal Violet, 500mL

Catalog Number: SS-041C-EU
Lot: K79106
Expiration Date: 2022, August 03

Characteristics	Specifications	Results
Appearance	Dark purple solution	Conforms
Optical density at 590 nm	0.700 – 0.820	0.810

Authorized by:

Whitney Lewis, Director of Quality

2021-06-17