



BIOMEDICAL SYSTEMS

370 West 1700 South
Logan, Utah 84321 USA
435 752 6011
800 453 2725
fax 435 752 4127
web www.elitechgroup.com

2022, March 09

Certificate of Analysis:

Gram Reagent C: Crystal Violet, 500mL

Catalog Number: SS-041C-EU
Lot: K09047
Expiration Date: 2023, May 10

Characteristics	Specifications	Results
Appearance	Dark purple solution	Conforms
Optical density at 590 nm	0.700 – 0.820	0.820

Authorized by:  03/09/2022

Lyndon Jay Smith, Quality Assurance Associate