

Negative Control

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name:_	Xpert [®]	C. difficile BT				
Cepheid Catalogue Pa	ırt No.:	GXCDIFFBT-CE-10)			
Kit Lot No.: 1000	737630			,		
Cartridge Lot No.: 35	201					
Kit Expiration Date:	2024-08-0	14				
Legal Manufacturer				Manufacturing Facility	Solna	○ Sunnyvale
Cepheid AB Röntgenvägen 5				Cepheid AB Röntgenvägen 5		
SE-171 54 Solna Sweden				SE-171 54 Solna Sweden		
Functional Testing		,				
Test Description	Acceptance Criteria				Test Result	
Positive Control	Toxigenic C.diff POS;Binary Toxin POS;027 PRESUMPTIVE POS					Passed

Toxigenic C.diff NEG;Binary Toxin NEG;027 NEG

Anton Andersson 2023-02-21
Signature of Quality Assurance Date

Name: Anton Andersson Engström

Title: QA Analyst

Passed