



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert® HCV Viral Load

Cepheid Catalogue Part No.: GXHCV-VL-CE-10

Kit Lot No.: 1001406411

Cartridge Lot No.: 38006

Kit Expiration Date: 2025-03-09

Legal Manufacturer

Cepheid AB
Röntgenvägen 5
SE-17154 Solna
Sweden

Manufacturing Facility

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna
Sweden

Solna Sunnyvale Newark

Functional Testing according to D51112 Rev: D

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
LOW Positive	HCV DETECTED: 1.50-1.90 log ₁₀ IU/mL	Passed
HIGH Positive	HCV DETECTED: 4.25-4.55 log ₁₀ IU/mL	Passed
Negative	HCV NOT DETECTED	Passed

20231227

Signature of Quality Assurance,

Date

Name: Robert Fiedler

Title: QA Analyst