

## • Certificate of Analysis

Product Name : **IVD Bacterial Test Standard**

Part No.: **8290190**

LOT No.: **6030324002**

Use by<sup>1)</sup>: **2026-01-31**

Date of Production: **2024-02-05**

Storage Conditions : **Store at -18 °C or below immediately on arrival.**

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
Verification of the calibration points by MALDI-TOF MS	Within a mass range of $\pm 900$ ppm	Passed
Relative intensity of reference protein No. 3 by MALDI-TOF MS	$0.20 \pm 0.05$	Passed
Relative intensity of reference protein No. 4 by MALDI-TOF MS	$0.22 \pm 0.05$	Passed
MALDI Biotyper identification	Classification Result Report lists " <i>Escherichia coli</i> " as best match.	Passed
	log(score) Value $\geq 2.2$ of the best match	Passed

### Remarks:

<sup>1)</sup> IVD BTS is shipped at ambient temperature but must be stored at  $-18^{\circ}\text{C}$  or below immediately on arrival. The expiry date on the package is valid for the enclosed IVD BTS when stored at  $-18^{\circ}\text{C}$  or below on arrival. Frozen, dissolved Bruker IVD BTS can be stored for up to 5 months at  $-18^{\circ}\text{C}$  or below. See instruction for use at [www.bruker.com/care](http://www.bruker.com/care).

Bruker Daltonics GmbH & Co. KG  
Fahrenheitstr. 4  
28359 Bremen  
Germany  
Phone +49-421-2205-0

QC release date: 2024-02-09

  
Quality Management