Version: 3.0

## **Declaration of Conformity**

for Nucleic acid sample preparation instrument IVD, automated Regulation (EU) 2017/746 of the European Parliament and of the Council of 5 April 2017 concerning In Vitro Diagnostic Medical Devices

The undersigned, under their sole responsibility, declares that the products described in this document meet the Council provisions that apply to them, and the CE Mark may be affixed.

General Product Name:	Nucleic acid sample preparation instrument IVD, automated	
Legal Manufacturer: (Name on Label)	AusDiagnostics Pty Ltd 290-292 Coward Street Mascot, NSW 2020 Australia	
SRN:	AU-MF-000020558	
Basic UDI-DI:	9343044060736APA	
Variants:	As per Appendix II (This document) – Product Listing/Schedule	
Intended Purpose:	For automated purification of nucleic acids from appropriate human specimen types.	
IVDR Classification:	Class A[Rule 5]	
Notified Body:	Not applicable for Class A	
CE Certificate:	Not applicable for Class A	
EU Authorised Representative:	Advena Limited. Tower Business Centre, 2 <sup>nd</sup> Flr., Tower Street, Swatar, BKR 4013 Malta.	
EU Authorised Representative SRN:	MT-AR-000000234	
IVDR Assessment Route:	For Class A: Issuing of the Declaration of Conformity in accordance with Article 17 after drawing up the technical documentation in Annexes II and III of the EU IVDR 2017/746.	

Name	_Joanna Haydar	Position	Senior Regulato	ry Affairs A	ssociate	
Signed	Light	Date	14-July-2022	Place	MASCOT	

Who is the natural and legal person with responsibility for the design, manufacture, packaging and labelling before the device is placed on the market under this manufacturer's name regardless of whether these operations are carried out by the manufacturer or on his behalf by a third party.

Version: 3.0

## Appendix I – Applicable Standards

This present declaration is also in conformity with the following European standards and Common Specifications (CS):

Standard/CS/Document Name	Description	
	Regulation (EU) 2017/746 of the European Parliament and of the	
2017/746	Council of 5 April 2017 concerning In Vitro Diagnostic Medical	
	Devices	
EN ISO 13485:2016+A11:2021	Medical Devices – Quality Management Systems – Requirements	
	for Regulatory Purposes	
EN ISO 14971:2019+A11:2021	Medical Devices – Application of Risk Management to Medical	
	Devices	
EN ISO 15223-1:2021	Medical devices. Symbols to be used with information to be	
	supplied by the manufacturer - General requirements	
EN ISO 20417:2021	Medical devices. Information to be supplied by the manufacturer	
	In vitro diagnostic medical devices — Information supplied by the	
EN ISO 18113-1:2011	manufacturer (labelling) — Part 1: Terms, definitions, and	
	general requirements	

## Appendix II – Product Listing/Schedule

Catalogue	UDI-DI	Device Description	EMDN
Number			Code
93100	09343044003578	MT-Prep 24 system	W02050106
93600	09343044004339	MT-Prep XL System	W02050106

## **Version History**

Version	Compiled by	Date	Description
3.0	Joanna Haydar	14/07/2022	93600- MT-Prep XL System and UDI-DI added
2.0	Axel Johannsson	12/05/2022	UDI-DI added
1.0	Axel Johannsson	26/04/2022	Original version