



www.pro-lab.co.uk

Certificate of analysis

Product Name: RABBIT COAGULASE PLASMA	Lot number: 240124a
Product Code: PL.850-10	Expiry date: 2027 02 28

Additional Information

Description: Rabbit plasma containing disodium EDTA as anticoagulant.	Pack size: 10ml Production date: 2024/02/19 Instructions for use: Go to www.pro-lab.co.uk for current version.	Storage conditions: 2-8°C Precautions: refer to SDS
---	---	--

Quality Control Testing

Raw Materials and Components:
All incoming chemicals, raw materials and components are quality controlled against predetermined specifications to ensure reproducibility.


Production:
Production is carried out against a Product Manufacturing Procedure incorporating Standard Operating Procedures and Good Manufacturing Practices.

Finished Product:
Lot to lot comparisons are carried out to ensure that the appearance of the material, as well as the specificity and sensitivity, remains constant.

Quality Assurance
Testing against S.aureus NCTC 12981 / ATCC 25923, S.aureus NCTC 6571 / ATCC 9144 and S.epidermidis NCTC 13360 / ATCC 12228 is carried out against predetermined parameters. The above lot passed all tests.

Stability and Shelf Life of Lyophilised Product
Recommended storage temperature:2-8°C. Storage at lower temperatures has no detrimental effect. Extended stability trials at 4°C, 25°C and 37°C have shown that the product has a minimum shelf life of 2 years at 4°C and at least one month at 25°C. Reconstituted material, using sterile distilled water, should not be stored longer than necessary and should never exceed the recommendations of the product literature.

Release

28/02/2024	
Date	Signed

T: 0151 353 1613
3 Bassendale Road, Bromborough,
Wirral, CH62 3QL, UK.
uksupport@pro-lab.com

