

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Clarithromycin CLR 0.016-256 µg/mL	051923018	19.05.2023	2026.05.18

Ref.

92048 – 920480 – 920481

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.12–0.5	0.19
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Agar	0.5 McFarland in broth	35 ± 2°C, ambient 20-24 h	4–16	6
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Agar	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	8–32	12
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, ambient 20-24 h	0.032–0.125	0.064
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	0.064–0.25	0.125

*Mueller Hinton II Agar (Cation Adjusted): Liofilchem ref. 10031; BD-BBL ref. 254032.

Batch Release

Approved

Date

08.06.2023

Signature

Quality Control

(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.