

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Micafungin MYC 0.002-32 µg/mL	021218092	12.02.2018	2021.02.11

Ref.

92182 – 921820 – 921821

Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents


Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida krusei</i> ATCC® 6258	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.032-0.25	0.064
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.25-2	0.38
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.004-0.032	0.006

Batch Release

Approved

Date

03.04.2018

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.