

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Oxacillin OX 1 µg	121119143	2022.12.10

Ref.

9036 – 9036/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	18–24	19
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	19–25*	21
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar with 5% Sheep Blood	0.5 McFarland	35 ± 1°C, 5% CO ₂ 20-24 h	≤ 12	10
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton Fastidious Agar	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-20 h	8–14*	10

*Established and validated by EUCAST

Batch Release

Approved

Date

16.12.2019

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.