

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Tetracycline TE 30 µg	070519078	2022.07.04

Ref.

9043 – 9043/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	18-25	19
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	24–30	24
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	23–31*	23
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Medium	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-18 h	14–22	15
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar with 5% Sheep Blood	0.5 McFarland	35 ± 1°C, 5% CO ₂ 20-24 h	27–31	28
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton Fastidious Agar	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-20 h	28–34*	28
<i>Neisseria gonorrhoeae</i> ATCC® 49226	GC Agar Base with 1% Defined Growth Suppl.	0.5 McFarland	36 ± 1°C, 5% CO ₂ 20-24 h	30–42	31
<i>Haemophilus influenzae</i> ATCC® 49766	Mueller Hinton Fastidious Agar	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-20 h	28–34*	29
<i>Campylobacter jejuni</i> ATCC® 33560	Mueller Hinton Fastidious Agar	0.5 McFarland	41 ± 1°C microaerobic 24 h	30–38*	33

*Established and validated by EUCAST

Batch Release

Approved

Date

15.07.2019

Signature

Quality Control
(D. Vitagliano)

Dario Vitagliano

The results reported were obtained at the time of release.