

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Tobramycin TOB 10 µg	072720059	2023.07.27

Ref.

9044 – 9044/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	18–26	20
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	19–29	23
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	20–26*	24
<i>Pseudomonas</i> <i>aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	20–26	23

*Established and validated by EUCAST

Batch Release

Approved

Date

30.07.2020

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.