

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Pefloxacin PEF 5 µg	042919058	2022.04.28

Ref.

9091 – 9091/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	25-33	29
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	26-32*	29

*Established and validated by EUCAST

Batch Release

Approved

Date

03.05.2019

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.