

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Nitrofurantoin F 100 µg	010819116	2022.01.07

Ref.

9158 – 9158/1

Antimicrobial Susceptibility Testing

Tested according to current EUCAST methodology


Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	17-23	20
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	17-23	19
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	18-24	23
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton Fastidious Agar	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-20 h	25-31	29

Batch Release

Approved

Date

17.01.2019

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.