

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Ceftaroline CPT 5 µg	092718043	2021.09.26

Ref.

9195 – 9195/1

Antimicrobial Susceptibility Testing

Tested according to current EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	24–30	25
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	24–30	27

Batch Release

Approved

Date

01.10.2018

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.