

Quality control certificate

Todd Hewitt Broth

N° Cat: 1236

N° Lote: 910012

Re-test Date: 10/2023



Formula in g/l

Bacteriological peptone	20	Dextrose	2
Disodium phosphate	0,4	Sodium carbonate	2,5
Sodium chloride	2	Heart infusion	3,1

Quality control

Solubility	Appareance	Color of the dehydrated medium	Color of the prepared medium	Final pH (25°)	Lote pH
w/o rests	Fine powder	Clear beige	Amber	7,8 ± 0,2	7,7

Microbiological test

Incubation conditions: (35±2 °C / 18-48 h)

Microrganisms	Specification
Neisseria meningitidis ATCC 13090	Good growth
Streptococcus pyogenes ATCC 19615	Good growth
Streptococcus pneumoniae ATCC 6305	Good growth
Streptococcus mitis ATCC 9895	Good growth

This certificate is an electronic copy of the certificate available in our laboratory and does not require signature.