

QUALITY CONTROL CERTIFICATE

SUCROSE Diagnostic Tablets

REF: 53811
LOT: 2001-1
Date of analysis: 2020-07
Date of expiry: 2024-07
Appearance: Plane tablet, 6 mm in diameter.
Complies.
Average weight: 90 mg.
Complies.

Performance:

Test strains:

Enterobacter cloacae ATCC 13047
Morganella morganii ATCC 25830
Providencia alcaligenes EF 6325
Proteus vulgaris ATCC 13315

Product complies with specification.

Rosco Diagnostica A/S



Peter Samuelson
Manager Quality Assurance